## **TECHNICAL DATA SHEET**

# THUNDER™ Phospho-SYK (Y525/Y526) TR-FRET Cell Signaling Assay Kit

CATALOG NUMBERS KIT-SYKP-100 (100 tests)

KIT-SYKP-100 (100 tests) KIT-SYKP-500 (500 tests) KIT-SYKP-2500 (2500 tests) KIT-SYKP-5000 (5000 tests) KIT-SYKP-10000 (10000 tests)



## bio**auxilium**

BETTER TOOLS. REAL DISCOVERIES.

#### **PRODUCT DESCRIPTION**

This assay kit measures intracellular levels of **phospho-SYK (Y525/Y526)** protein in cell lysates using a simple, rapid and sensitive immunoassay based on the homogeneous (no-wash) THUNDER<sup>™</sup> TR-FRET technology. The kit is compatible with both adherent and suspension cells.

## SPECIFICITY

This assay kit contains two specific and selective antibodies, one that recognizes **SYK** phosphorylated at **Tyr525/Tyr526**, and another that recognizes an invariant epitope of **SYK**.

Store at -80°C

For research use only.

Not for use in diagnostic procedures.

## SPECIES REACTIVITY

Human (Swiss-Prot Acc.: P43405; Entrez-Gene Id: 6850).

Other species should be tested on a case-by-case basis.

## **TR-FRET ASSAY PRINCIPLE**

The Phospho-SYK (Y525/Y526) assay kit is a homogeneous timeresolved Förster resonance energy transfer (TR-FRET) sandwich immunoassay (Figure 1). The THUNDER<sup>™</sup> Cell Signaling assay workflow consists of 3 steps (Figure 2). Following cell treatment, cells are first lysed with the specific Lysis Buffer provided in the kit. Then Phospho-SYK (Y525/Y526) in the cell lysates is detected with a pair of fluorophore-labeled antibodies in a simple "add-incubatemeasure" format (single-step reagent addition; no wash steps). One antibody is labeled with a donor fluorophore (Europium chelate: Eu-Abl) and the second with a far-red acceptor fluorophore (FR-Ab2). The binding of the two labeled antibodies to distinct epitopes on the target protein takes place in solution and brings the two dyes into close proximity. Excitation of the donor Europium chelate molecules with a flash lamp (320 or 340 nm) or a laser (337 nm) triggers a FRET from the donor to the acceptor molecules, which in turn emit a TR-FRET signal at 665 nm. Residual energy from the Eu chelate generates light at 615 nm. The signal at 665 nm is proportional to the concentration of Phospho-SYK (Y525/Y526) in the cell lysate. Data can be expressed as either the signal at 665 nm or the 665 nm/615 nm ratio.

STEP 1	STEP 2	STEP 3	
Cell treatment	Cell lysis	Protein detection	
Seed cells in culture plate     Add media +/- compound     Incubate for optimized time	Adherent cells: remove media and add 1X Lysis Buffer Suspension cells: add 5X Lysis Buffer Incubate for <b>30 min</b>	<ul> <li>Transfer lysate         <ul> <li>(15 μL) to detection             plate</li> <li>Add 4X Antibody             Mix (5 μL)             lucubate for 2 h</li> </ul> </li> </ul>	
-,		• Read TR-FRET signal	

Figure 2 Assay workflow using the 2-plate (transfer) protocol.




Figure 1 Schematic representation of the TR-FRET cell signaling assay principle.

KIT COMPONENTS	100 points*	500 points*
Eu-labeled phospho-SYK (Y525/Y526) antibody (Eu-Ab1)	5μL	25 µL
Acceptor-labeled phospho-SYK (Y525/Y526) antibody (FR-Ab2)	20 µL	100 µL
Lysis Buffer 3 (5X)	lmL	5 mL
Detection Buffer (10X)	50 µL	250 µL
Positive control cell lysate	100 µL	500 µL
Phosphatase Inhibitor Cocktail (100X)	50 µL	250 µL

\* The number of assay points is based on an assay volume of 20 μL in halfarea 96-well or low-volume 384-well assay plates using the kit components at the recommended concentrations (refer to the User Manual).

## **TECHNICAL DATA SHEET** Phospho-SYK (Y525/Y526)


## **VALIDATION DATA**

IN RAJI CELLS


This assay kit has been validated for the relative quantification of phospho-SYK (Y525/Y526) in Raji and U-937 cell lysates using the 2-plate assay protocol.

- Non-adherent cells were cultured in RPMI+10% FBS before being centrifuged and resuspended at the desired density in RPMI without serum.
- Following cell treatment, cells were lysed with the 5X Lysis Buffer **3** supplemented with the 100X Phosphatase Inhibitor Cocktail diluted at 5X

STIMULATION OF PHOSPHO-SYK (Y525/Y526)



Raji cells (100,000 cells/well; in triplicate) were incubated with serial dilutions of pervanadate for 30 min at 37°C. Data show that treatment of Raji cells with pervanadate stimulates phosphorylation of SYK at Y525/Y526.



20

## Z'-FACTOR DETERMINATION

10

Raji cells (100,000 cells/well) were incubated without or with 150  $\mu$ M of pervanadate 30 min at 37°C. The Z' factor value was determined using a total of 48 wells for each treatment group. The Z'-factor value of 0.82 indicates that the assay is robust and suitable for HTS.

Well #


30

40

50


- Following a **30-min** incubation at room temperature (RT) on an orbital shaker (400 rpm), lysates (15 µL) were then transferred to a 384-well assay plate followed by addition of the labeled antibodies Eu-Ab1 and FR-Ab2 (5 µL) for detection of phospho-SYK (Y525/Y526).
- The plate was incubated at RT for **2 hours** and the TR-FRET signal was recorded at 665 and 615 nm (EnVision®; laser excitation).

#### INHIBITION OF PHOSPHO-SYK (Y525/Y526) IN U-937 CELLS



U-937 cells (50,000 to 300,000 cells/well; in triplicate) were incubated with serial dilutions of BIIB-057 for 60 min at 37°C. Cells were then stimulated with 200 µM of pervanadate for 30 min at 37°C. Data show that treatment of U-937 cells with BIIB-057 inhibits phosphorylation of SYK at Y525/Y526 by pervanadate.

### **U-937 CONTROL LYSATE TITRATION** (OC TEST)



Quality Control: the Phospho-SYK (Y525/Y526) assay kit is routinely tested against pervanadate-treated U-937 lysates. U-937 were cultured, centrifuged, resuspended at 10 million cells/mL and stimulated with 200 µM of pervanadate for 30 min at 37°C. Following cell lysis. lysates were serially diluted with 1X Lysis Buffer 3 and tested in triplicate. Data show a linear relationship between lysates prediluted at 1:4 or more and TR-FRET ratio values.



0

#### FOR MORE INFORMATION ON DEVELOPING AND OPTIMIZING TR-FRET CELL SIGNALING ASSAYS, CONSULT THE USER MANUAL.